



**Bureau of Laboratory Quality Standards  
Ministry of Public Health**

This is to certify that

**The Laboratory of**

**Calibration and Testing Equipment Services**

**Technology Promotion Association ( Thailand – Japan )**

**534/4 Soi Pattanakarn 18 , Pattanakarn Road ,  
Suanluang , Bangkok 10250 , Thailand.**

has been accepted as an  
accredited laboratory complying with the ISO/IEC 17025 : 2017  
and the requirements of the Bureau of Laboratory Quality Standards

The laboratory has been accredited for specific tests  
listed in the scope within the field of

**Medical Devices Testing**

*Patravee Soisangwan*

( Dr. Patravee Soisangwan )

Director of Bureau of Laboratory Quality Standards

Date of Accreditation : 25 June 2020

Valid Until : 24 June 2022

Accreditation Number 1248/57

**The Laboratory of Calibration and Testing Equipment Services , Technology Promotion Association ( Thailand-Japan ) has been accepted as accredited laboratory in the field of medical devices testing for the following scopes.**

No.	Type of Sample	Test	Method
1.	Non-invasive mechanical sphygmomanometers	1. Maximum permissible errors of the cuff pressure indication	- International Organization of Legal Metrology. OIML R16-1 : 2002 (E) - ISO 81060-1 : 2007 ( E )
		2. Air leakage	
		3. Hysteresis error	
2.	Non-invasive automated sphygmomanometers	4. Maximum permissible errors of the cuff pressure indication	- International Organization of Legal Metrology. OIML R16-2 : 2002 (E) - ISO 81060-2-30 : 2009 ( E )
		5. Air leakage	